

Ben G. Almond • Vice President, Regulatory Affairs • phone 202,419,3020 • fax 202,419,3047

March 6, 2003

RECEIVED

Ms Marlene H Dortch Secretary Federal Communications Commission 445 12<sup>th</sup> Street, SW, Room TW-A325 Washington, DC 20554

MAR = 6 2003

DEFICE OF THE SECRETARY

KE Section 68.4(a) of the Commission's Rules Governing Hearing Aid Compatibility Telephone WT Docket No 01-309 **EX PARTE** 

Dear Ms Dortch:

On March 5, 2003, representatives of Siemens and Cingular Wireless met in a joint meeting with members of the Wireless Telecommunications Bureau, Consumer and Governmental Affairs Bureau and the Office of Engineering and Technology to discuss issues related to the referenced proceeding

Two documents were used for discussion purposes, the attached document and the ex parte document previously submitted by Siemens and Cingular Wireless on January 22, 2003 Please associate this notification and accompanying material with the referenced docket proceeding.

The list of attendees for the meeting is attached. If there are any questions concerning this matter, please contact the undersigned.

Sincerely,

Ben G Almond

Vice President-Federal Regulatory Affairs

Attachments

Cc Joel Taubenblatt

Mindy Littell

Pat Forster

Greg Guice

Janet L Sievert

Jerry Stanshine

OH/

### Siemens/Cingular Wireless. Ex Parte-Docket 01-309 March 5, 2003

Attendees:

**FCC** Siemens


Ross Vincenti Joel Taubenblatt-WTB Mindy Littell-WTB Eleanor Kerr

Pat Forster-WTB Stephen Berger (consultant)

Greg Guice-WTB Janet Sievert-DRO

Cingular Wireless Jerry Stanshine-OET

Susan Palmer Ben Almond



# Hearing Aid Compatibility

Testing & Technical Update

**Cingular Wireless Siemens** 

Ex Parte - WT Docket 07-309

# Overview

Results of Siemens hearing aid and handset testing

T-Coil and Functional Equivalency ANSI C63. 19 Background

Cingular/Siemens Earlier Ex Parte Recommendations (Chart)

- Technical Incubator and Steering Committee
- Communication/Education

# **Technical Overview**

The Siemens handset and hearing aids performed well together

- ANSI C63.19 accurately predicted performance
- To achieve our recommended T-Coil performance flexible design options are needed


# **Results of Siemens Testing**

Siemens handsets were tested and achieved the U3 & U4 category per ANSI C63.19

Siemens Triano S hearing aid also achieved the U4 category for RF immunity

When tested together the Triano S hearing aid had no audible interference near the handsets in microphone mode, but some interference in T-Coil mode.

# Siemens S46 Handset plots – E-Field Plot



□ 54-56
■ 52-54
□ 50-52
□ 48-50
□ 44-46
□ 42-44

### **ANSI C63.19 System Classification**

	Articulation	l
System Classification	Index	U Category Sum
	Al	Sum of Hearing Aid (U Category) + Telephone (U Category)
Usable	0.3	= 4
Normal Use	0.5	= 5
Excellent Performance	0.7	>= 6

- •The combination of the U category of the phone and the hearing aid must equal 5 or greater to achieve the recommended performance level.
- The most appropriate forum to set specific limits is a technical standards committee with both industries represented.

# **Next Step User Testing**

Lab testing with hearing aids users will be conducted with Gallaudet University

Field testing will then be conducted with actual hearing aid users

Target Date for Completion: Late Spring

A report with analysis and summary of the results will be made available and is expected to be useful in developing consumer guidance

# Flexible Design Options Should Be Considered

T-Coils may not always be the best answer

When given a technical choice inductive loop (T-Coil) systems sell the least even though they are the cheapest

Increasing level of electronic saturation in work and public environments make T-Coil use impossible in many locations.

Only a small percentage of hearing aid users use T-Coils

# Flexible Design Options Should Be Considered

Form factors, especially in small phones, impact ability to build in T-coil antenna

A dynamic speaker will produce a T-Coil signal that may pass the current Part 68 level but be too low for many users

A more effective solution requires much higher signal levels and so a separate T-Coil antenna and special circuitry

Other forms of coupling should be allowed if functionally equivalent

Solutions should be identified and tested in a Technical Incubator

# **ANSI C63.19 Questions**

Posed by the FCC and others

Why conduct testing in analog mode?

Analog phones can interfere with hearing aids, contrary to common belief. However, when they interfere the symptom is not an audible 'buzz' but rather changes to the hearing aid gain. The analog signal can interfere with the gain setting of some hearing aids.

# **ANSI C63.19 Questions**

Posed by the FCC and others

What are issues that can make the results of user testing appear inconsistent with ANSI tests? In many cases user testing is performed without proper controls. Common problems are not knowing the immunity of the user's hearing aid or the transmission level of the phone.

# ANSI C63.19 Questions

Posed by the FCC and others

How can the results of hearing aid and handset testing help consumers?

can be guided to effective solutions are properly matched, consumers with a high degree of confidence. When hearing aids and handsats

# Cingular/Siemens Recommendations

Technical Focus – unbiased assessment Technical Incubator with engineers specializing in hearing and wireless technologies that develop and test solutions

- Independent Steering Committee, not Federal Advisory Committee, to guide efforts
- FCC/FDA must actively participate

## **Education and Outreach**

In addition to guiding the Technical Incubator, the Steering Committee can address non-technical issue

- Effective use of handsetsDetermining what combinations work can be complex.
- Support by audiologist and other relevant professionals may be required.

The hearing aid and wireless industry, consumers and hearing health professionals must work together to simplify product selection and provide additional assistance.

# **Summary of Technical Issues**

Testing demonstrated that Siemens hearing aid designs effectively mitigated interference The ANSI C63.19 standard is effective.

Minor variations in user testing can impact apparent outcome.

Variables must be monitored carefully to test effectively.

- The standard should be updated to address the current state of technology.
- Our recommended level of T-Coil performance is feasible with an external accessory

# A Cooperative and Comprehensive Approach is Needed

Cingular/Siemens believe that both technical and user issues must be addressed
Both wireless and hearing aid manufacturers must work toward a cooperative solution-based approach, endorsed by both the FCC and FDA.

### **Steps should include:**

- An unbiased technical assessment of the issue
- A Technical Incubator with a steering committee with active participation and monitoring by both agencies
- Development of easy and consistent information on product selection and usage